

No.ICFOSS/59/2023-TC4(IT)Part(1)

14/03/2024

QUOTATION NOTICE

Competitive quotations are invited from qualified firms/agencies for Dissolved Oxygen Sensor. Specification details are as follows

Sl.No	Description	Quantity
1	<p><u>Dissolved Oxygen Sensor</u> Dissolved Oxygen Prob: Type: Galvanic Probe, Detection Range:0~20mg/L, Response Time: Up to 98% full response, within 90 seconds (25°C), Pressure Range: 0~50PSI, Electrode Service Life: 1 year (normal use), Maintenance Period: Membrane Cap Replacement Period: 1~2 months (in muddy water); 4~5 months (in clean water) Filling Solution Replacement Period: Once every month, Cable Length:2 meters, Probe Connector: BNC</p> <p>Signal Converter Boar: Operating Voltage: 3.3~5.5V, Output Signal: 0~3.0V, Cable Connector: BNC, Signal Connector: Gravity Analog Interface (PH2.0-3P), Dimension: 42mm * 32mm</p> <p>Package Include: Galvanic Dissolved Oxygen Probe with Membrane Cap x1, Spare Membrane Cap x1, Signal Converter Board x1, Gravity Analog Sensor Cable x1, Waterproof Gasket x2, Screw Cap for BNC Connector x1, Plastic Dropper x2</p>	3

Rate should be quoted in gross amount including GST. Sealed quotation shall be reached this office on or before 4pm on 21/03/2024 to the undersigned marked properly on the cover "Quotation for Dissolved Oxygen Sensor"

Quotation will be opened on 21/03/2024 at 4.30pm



Chitra
Chitra M.S
Secretary & Registrar
ICFOSS